SIEMENS

ARCADIS Varic

SP

Installation Instructions

System

Replacing the US Keyboard

© Siemens AG 2007
The reproduction, transmission or use of this document or its contents is not permitted without express written authority. Offenders will be liable for damages. All rights, including rights created by patent grant or registration of a utility model or design, are reserved.

Print No.: SPR2-310.812.06.01.02
Replaces: n.a.

English
Doc. Gen. Date: 07.07

Document revision level

The document corresponds to the version/revision level effective at the time of system delivery. Revisions to hardcopy documentation are not automatically distributed.

Please contact your local Siemens office to order current revision levels.

Disclaimer

The installation and service of equipment described herein is to be performed by qualified personnel who are employed by Siemens or one of its affiliates or who are otherwise authorized by Siemens or one of its affiliates to provide such services.

Assemblers and other persons who are not employed by or otherwise directly affiliated with or authorized by Siemens or one of its affiliates are directed to contact one of the local offices of Siemens or one of its affiliates before attempting installation or service procedures.

Table of Contents 3

1 Requirements 4

. 4
General 4
Required Parts 4
Documents required 4
Required tools 4
Required aids 4
Contents of the spare parts kit for Varic/Orbic operating consoles 8081536 5

2 Required Steps 6

. 6
Opening the monitor cart 6
Replacing the keyboard 6
Functional Test 6
Closing the covers, safety checks 6
Selecting the required CD-ROM 6
Selecting the required language for the "important supplemental information for the operating instructions" 6
Printing the "important supplemental information for the operating instructions" file . 7
Completing the operating instructions. 7
Customer information 7

3 Changes to previous version 8

General

Since approx. June 2007, a new keyboard (text instead of icons, material no. 8082070) has been installed in monitor carts intended for delivery to the USA.

Beginning with the following system serial numbers, the new monitor cart keyboard will be delivered with systems sent to the USA or when this keyboard is explicitly ordered by the customer:

- ARCADIS Varic beginning with serial number 10759
- ARCADIS Orbic beginning with serial number 20358
- ARCADIS Avantic beginning with serial number 31168

The old design of the monitor cart keyboard (icons) with material number 8079514 will no longer be delivered as a spare part.

When replacing the old monitor cart keyboard 8079514 with the new keyboard 8082070, the operating instructions on-site have to be supplemented.

For this reason, kit 8081536 contains two CD-ROMs with the corresponding files in PDF format.

Required Parts

- Monitor cart keyboard 8082070 (text)
- Spare parts kit for Varic/Orbic operating consoles 8081536

Documents required

- Service documentation for the system (ARCADIS Varic or ARCADIS Orbic)
- System manual for the system

Required tools

- Standard service tool kit

Required aids

- Service PC with CD-ROM drive and WINDOWS XP operating system, installed PDF reader, and printer.
- Installed PDF reader on the service PC
- Printer to print out the supplemental information to add to the operating instructions.
- Standard service tool kit

Contents of the spare parts kit for Varic/Orbic operating consoles 8081536

- CD-ROM “Important additional information - ARCADIS Varic Multilingual”

  The "important supplemental information for the operating instructions" in all available languages is stored on the CD as PDF files.

  ➪ Print number SPR2-310.601.02.01.02

- CD-ROM “Important additional information - ARCADIS Orbic Multilingual”

  The "important supplemental information for the operating instructions" in all available languages is stored on the CD as PDF files.

  ➪ Print number SPR2-320.601.02.01.02

- These installation instructions SPR2-310.812.06.01.02

NOTE

Adhere to the safety information in the corresponding system service documentation.

The opening and closing of covers is described in the corresponding system service documentation.

Opening the monitor cart

Open the monitor cart as described in the system service documentation.

Replacing the keyboard

Replace the existing keyboard with the new keyboard.

Functional Test

Perform a functional test of the system.

Closing the covers, safety checks

Close all covers and perform the required safety checks (protective conductor test) in accordance with the system service documentation.

Selecting the required CD-ROM

Remove the required CD-ROM (ARCADIS Varic or ARCADIS Orbic) from kit 8081536 and place it in the CD-ROM drive of your service PC.

Selecting the required language for the "important supplemental information for the operating instructions"

Select the file required for printing in the language of the system operating instructions available on-site. Use the Explorer of the operating system on your service PC.

NOTE

The last two digits of the file name (before ".pdf") correspond to the language code of the stored document. Refer to the "_Readme.pdf" for the languages corresponding to the language codes.

The “Readme.txt” file contains the language codes and corresponding languages for the "important supplemental information for the operating instructions".

Printing the "important supplemental information for the operating instructions" file

Print the selected file.

Completing the operating instructions

Attach the printed pages to the on-site operating instructions.

Customer information

Inform the customer of the modified layout of the monitor cart keyboard.

Inform the customer on the "important supplemental information for the operating instructions" that has been added.

Provide the customer with the operating instructions and CD-ROM.

8 Changes to previous version

Not applicable, new issue.